



Barcode Label Placement Instructions

GX-281 Liquid Handler

Rack Heightener

If you are using a rack heightener with a Code 200-series rack on the GX-281 Liquid Handler, use the following instructions.

1. Locate the barcode label included with the rack heightener and remove the label from the backing.
2. Place the rack heightener label between the two lines on the rack barcode label. The text should be facing out (the same orientation as the rack barcode label).

